



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 1, 2010 INITIALS: DEC

OWNER: WASHINGTON STATE PATROL

DISTRICT: 6 ADDRESS: I-90W Bullfrog

LOCATION CleElum 53 WB S

COUNTY: Kittitas

MAKE: GSE

TYPE Trk Scale

ACCURACY CLASS: III L

MODEL: 651

ID. NO: 202173

CAPACITY: 200000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: Digital

PLATFORM 12x20

Dual South

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20000	20000	0	SEC 1	20000	20000	0
CENTER	20000	20000	0	CENTER	20000	20000	0
SEC 2	20000	20000	0	SEC 2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer

Date: July 1, 2010

Print name: David Cromer

State of Washington

County of Shurston

Signed and sworn to (or affirmed before
me on July 1, 2010 by

Brenda A Biscay

(Signature of notary public)

Brenda A Biscay

Printed Name of Notary

My appointment expires 3.12.12

